



## CERTIFICATE OF ANALYSIS

|                 |   |                        |             |
|-----------------|---|------------------------|-------------|
| <b>CONTACT:</b> | KENNETH CHEUNG  | <b>WORK ORDER:</b>     | HK1775666   |
| <b>CLIENT:</b>  | OPTIWAY (ASIA) CO. LTD  | <b>SUB-BATCH:</b>      | 1           |
| <b>ADDRESS:</b> | FLAT B, 14/F,   | <b>LABORATORY:</b>     | HONG KONG   |
| <b>ADDRESS:</b> | SHUI WING INDUSTRIAL BUILDING,<br>12-22 TAI YUEN STREET KWAI CHUNG,<br>N.T. , HONG KONG | <b>DATE RECEIVED:</b>  | 27-Oct-2017 |
|                 |   | <b>DATE OF ISSUE:</b>  | 27-Nov-2017 |
|                 |   | <b>SAMPLE TYPE:</b>    | LIQUID      |
|                 |   | <b>NO. OF SAMPLES:</b> | 1           |

### COMMENTS

Sample(s) were picked up from client by ALS Technichem (HK) staff in an ambient condition.

Sample(s) analysed and reported on an as received basis.

Testing period: 27-Oct-2017 to 27-Nov-2017

#### **Method Reference:**

BS EN 1276:2009 Incorporating Corrigendum August 2010, Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic, and institutional areas - Test method and requirements (phase 2. step 1)

Appearance of the product: Transparent and colorless liquid

Active substance(s) and their concentration(s): Unknown

Product diluent recommended by the manufacturer for use: Not applicable

Method: Membrane filtration

#### **Conclusion:**

The bactericidal activity of the submitted sample at original concentration showed greater than 5 lg reduction (lg R) on *Candida albicans* under dirty condition which standing at 20 °C with 15 minutes contact time, according to the BS EN 1276 standard.

### SAMPLE DETAILS

| ALS Lab ID    | Product Name                   |
|---------------|--------------------------------|
| HK1775666-001 | NICE - CRYSTAL SANITIZER コレスゴ! |

### NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Mr Luk Hon Yin, Henry  
Microbiologist - Microbiology

# CERTIFICATE OF ANALYSIS



**WORK ORDER :** HK1775666  
**SUB-BATCH :** 1  
**CLIENT :** OPTIWAY (ASIA) CO. LTD  
**DATE OF ISSUE :** 27-Nov-2017

**Results:**

| Test Organism                         | Validation test                          |                              |                        | Test suspension       |         |      | Result                |                   |       |                    |
|---------------------------------------|--|------------------------------|------------------------|-----------------------|---------|------|-----------------------|-------------------|-------|--------------------|
|                                       | Validation suspension (NV <sub>0</sub> ) | Experimenta l conditions (A) | Filtration control (B) | Method validation (C) | N       | lg N | N <sub>0</sub> = N/10 | lg N <sub>0</sub> |       |                    |
| <i>Candida albicans</i><br>ATCC 10231 | 40                                       | 39                           | 44                     | 41                    | 4.6E+08 | 8.66 | 4.6E+07               | 7.66              | Na    | contact time (min) |
|                                       |  |                              |                        |                       |         |      |                       |                   | lg Na | 15                 |
|                                       |  |                              |                        |                       |         |      |                       |                   | lg R  | 3.0E+01            |
|                                       |  |                              |                        |                       |         |      |                       |                   | R%    | 1.48               |
|                                       |  |                              |                        |                       |         |      |                       |                   |       | 6.19               |
|                                       |  |                              |                        |                       |         |      |                       |                   |       | 99.99%             |

NV<sub>0</sub> = number of cfu/mL in the mixtures "A", "B" and "C" at the beginning of the contact time (time 0)

A = number of cfu/mL of the experimental conditions control

B = number of cfu/mL of the filtration control control

C = number of cfu/mL of the method validation

N = number of cfu/mL in the test suspension

N<sub>0</sub> = number of cfu/mL in the test mixture at the beginning of the contact time (time 0)

Na = number of cfu/ml in the test mixture at the end of the contact time and before membrane filtration

lg R = reduction expressed in logarithm

R% = reduction expressed in percentage

# CERTIFICATE OF ANALYSIS



WORK ORDER : HK1775666  
SUB-BATCH : 1  
CLIENT : OPTIWAY (ASIA) CO. LTD  
DATE OF ISSUE : 27-Nov-2017

## **Experimental conditions:**

Product diluent used during the test: sterile tryptone sodium chloride solution  
Contact time :  $t = 15 \text{ mins} \pm 10 \text{ seconds}$   
Test temperature:  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$   
Interfering substance: 3g/l bovine albumin (dirty condition)  
Temperature of incubation for *Candida albicans*:  $25^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$

## **Basic limits:**

For each organism check that:

- a) N is between  $1.5 \times 10^8$  and  $5.0 \times 10^8$ ,  $N_0$  is between  $1.5 \times 10^7$  and  $5.0 \times 10^7$
- b)  $N_{V0}$  is between 30 and 160
- c) A, B, C are equal to or greater than  $0.5 \times N_{V0}$

## **Remarks regarding the results:**

All controls and validation were within the basic limits.

No precipitate during the test procedure (test mixtures were homogeneous).